




Occlutech UNI Occluder

Compatibility chart & ordering information



The Procedure Pack
INCLUDES
The Accessory Medical Devices
Flex-Pusher II



Ø LA DISC: LEFT ATRIAL DISC DIAMETER
Ø RA DISC: RIGHT ATRIAL DISC DIAMETER

Procedure Pack Art. No.	AS Defect Size (D) [mm]	Smallest Dist.* [mm]	Ø LA Disc [mm]	Ø RA Disc [mm]	Flex-Pusher II** Size / Color	ODSv1*** Size [F]	ODS III*** Size [F]
16UNI17F	D ≤ 8.4	8.5 – 11.9	17	17	51FP100 ●	7	7
16UNI24F	D ≤ 11.9	12 – 13.9	24	24	51FP120 ●	9	9
16UNI28F	D ≤ 13.9	14 – 16.4	28.5	28.5	51FP120 ●	9	9
16UNI33F	D ≤ 16.4	A ≥ 16.5	33	33	51FP160 ●	11	11
16UNI40F	D ≤ 19.5	A ≥ 20	40	40	51FP160 ●	12	12

Availability subject to regulatory approval. Product data may be subject to change. Contact your Occlutech representative for information.

* Smallest distance between defect and surrounding structures (aortic roof / opening of V. Cava Superior).

** Color defines Pusher color.

*** Figulla Flex II UNI occluder is compatible with Occlutech Delivery Set, ODSv1 (Article no.: 51DSXXX, 45°) and Occlutech Delivery Set III, ODS III (Article no.: 98DSXXX, 45°).



Occlutech International AB
Landskronavägen 2, SE-252 32 Helsingborg, Sweden
order@occlutech.com, www.occlutech.com

